

## Certificate

---

**Manufacturer** ResMed Paris  
240 rue de la Motte  
77550 Moissy-Cramayel  
France

---

**Products** Eole 3 XLS  
VS Integra, VS Ultra, VS III  
Elisée 150, Elisée 250, Elisée 350

---

We hereby testify that, to our knowledge, using our medical ventilators operating on their internal batteries in an airplane does not result in any specific risk.

Regarding their electromagnetic compatibility, the devices comply with EN 60601-1-2 standard specifications.

Concerning the used batteries, the technology reveals no specific risk and the batteries contain neither acids nor corrosive agents (dry batteries).

For Elisée ventilators equipped with Li-Ion internal batteries, we certify that the battery pack watt hour rating is less than 100 Wh (less than 8 Wh per Cell), and therefore Elisée ventilators are not submitted to the recommendations regarding air transportation of dangerous goods ST/SG/AC.10/1 and ST/SG/AC.10/11.

Moissy-Cramayel, January 4<sup>th</sup> 2010



Nathalie Jubert  
Director quality & regulatory  
ResMed Paris