

Mölnlycke, 2014-May-26

Breas Ventilator Vivo 50 - FAA Compliance

This FAA compliance letter is valid for Vivo 50s that has been produced from July 08 2013 with serial numbers starting from J270762.

The U.S Department of Transportation (DOT) Final Rule, "Non-discrimination on the Basis of Disability in Air Travel" (73 FR 27614 which updates Title 14 CFR Part 382), effective May 13, 2009 provides important new requirements for the accommodation of passengers with respiratory assistive devices (Ventilators, Respirators and CPAP machines).

In line with these requirements, respiratory assistive devices may be used on-board an aircraft, without further testing by the carrier, provided they have been tested for Electromagnetic Compatibility (EMC) in accordance with the current version of RTCA/DO-160, Section 21, Category M.

Breas has completed testing for the Vivo 50 Ventilator.

The Vivo 50 Ventilator complies with RTCA/DO-160G, Section 21, Category M and is FAA (Federal Aviation Administration) compliant.



Ann-Sofie Svensson

Vice President Quality and Regulatory Affairs

Breas Medical AB